



Characterisation of CR-39 Detectors at the CERF Reference Facility at CERN



M. Caresana^a, M. Ferrarini^{a,c}, A. Sashala Naik^{a,b}, A. Parravicini^b

^aPolitecnico di Milano, CESNEF, Dipartimento di Energia, via Ponzio 34/3, 20133 Milano, Italy

^bMi.am srl, via De Amicis 5, 29029 Fabiano di Rivergaro (PC), Italy

^cCNAO Fondazione, Via Privata Campeggi, 27100 Pavia, Italy



This research project has been partially supported by the Marie Curie Initial Training Network Fellowship of the European Community's Seventh Framework Programme under Grant Agreement PITN-GA-4 2011-289198-ARDENT

1) Introduction

The CERN-EU high-energy Reference Field (CERF) is a calibration facility dedicated to detector testing, where above its concrete roof, a neutron field similar to the one encountered at commercial flight altitudes of 10 to 20 km can be found. The CR-39 detectors were irradiated at 2 different positions (CT5 and CT12).




Fig 1. Complex neutron field on the concrete Roof-Shield situated right above the irradiated target

2) Materials





Fig 2. Layout of the passive detection system at the position CT 12 which is right on top of the irradiated target (neutron source).

| | | | |
|------|------|-------|-------|
| CT 1 | CT 5 | CT 9 | CT 13 |
| CT 2 | CT 6 | CT 10 | CT 14 |
| CT 3 | CT 7 | CT 11 | CT 15 |
| CT 4 | CT 8 | CT 12 | CT 16 |

TARGET

Fig 3. Top view of the concrete roof of the CERF facility showing the positions CT 5 and CT 12 of the neutron detectors.

3) Track Analysis and LET determination



(a) Politrack™ SSNTD automatic reader




(b) Raw image captured with Politrack



(d) Measurement of D and d



(c) On line image analysis



(e) LET distribution (the suffix nc indicates that it is a quantity measured with Nuclear track detector CR39)

5) Results




Fig 4. LET distribution measured for the position CT 5




Fig 5. LET distribution measured for the position CT12

| Position | CT 5 | CT 12 | Uncertainty (2σ) |
|-------------------------|------|-------|------------------|
| Dose in CR-39 (mSv) | 3.8 | 4.1 | ± 1.4 |
| Reference Dose (mSv) | 5.17 | 4.85 | ± 0.5 |
| Dose < 100 keV/μm (mSv) | 2.0 | 2.2 | |
| Dose > 100 keV/μm (mSv) | 1.7 | 2.1 | |

Table 2. Dose calculation from measured LET spectra taking into account the mean sensitivity measured from calibration campaigns illustrated in Table 3.

The results obtained for the dose measured at position CT5 and CT12, in Table 2, are based on the average sensitivity measured in calibration campaigns using monoenergetic neutron beams (Table 3, to be published in NEUDOS 12 proceedings) and confirms the suitability of this detection system for dosimetry in workplace fields.

| Beam | Measured Dose (mSv) | Reference Dose(mSv) | Sensitivity |
|----------------------------|---------------------|---------------------|--------------------|
| PTB 565 KeV | 1.79 | 3.67 | 0.49 |
| PTB 8 MeV | 1.75 | 4.90 | 0.36 |
| PTB 14 MeV | 3.49 | 6.90 | 0.51 |
| PTB 19 MeV | 1.84 | 2.90 | 0.64 |
| iThemba 66 MeV 0° | 2.38 | 4.44 | 0.54 |
| iThemba 66 MeV 16° | 1.72 | 3.20 | 0.54 |
| iThemba 100 MeV 0° | 1.52 | 2.36 | 0.64 |
| iThemba 100 MeV 16° | 1.75 | 2.83 | 0.62 |
| Average Sensitivity | | | 0.54 ± 0.09 |

Table 3. Average sensitivity calculated from calibration experiments done in monoenergetic neutron beam ranging from 0.5 MeV to 100 MeV.

4) Repeatability of measurements

The χ^2 is the statistical check to control that the tracks are Poisson distributed on the detector surface. The measurements follow closely the expected χ^2 value of 1.

| Detectors at CT 5 | χ^2 | Dose (H*10) in mSv | Detectors at CT 12 | χ^2 | Dose (H*10) in mSv |
|-------------------|-------------|--------------------|--------------------|-------------|--------------------|
| 1810 | 1.03 | 1.93 | 1818 | 0.98 | 2.26 |
| 1811 | 1.00 | 1.98 | 1829 | 0.91 | 2.22 |
| 1839 | 1.07 | 2.34 | 1840 | 0.97 | 2.33 |
| 1846 | 1.05 | 2.11 | 1856 | 1.03 | 2.18 |
| 1850 | 0.96 | 2.28 | 1868 | 1.08 | 2.19 |
| 1855 | 0.95 | 1.82 | 1877 | 0.97 | 2.18 |
| 1871 | 0.92 | 2.05 | 1895 | 0.95 | 1.98 |
| 1893 | 0.97 | 1.86 | 1906 | 1.06 | 2.40 |
| Mean Dose | 2.07 | | Mean Dose | 2.22 | |
| St. Dev. % | 8.90 | | St. Dev. % | 5.60 | |
| Ref. Dose | 5.17 | | Ref. Dose | 4.85 | |

Table 1. Dose measurement for each detector exposed at CERF at CT 5 and CT12

6) Conclusion

➤ The dosimeter, that was calibrated in monoenergetic fields of various energies, at the PTB and Ithemba labs, proved effective in measuring the dose in a workplace field having a significant high-energy component.